

When taking a reading, hold DipCell comparator so that you are looking above the horizon with sunlight coming over your shoulder.

Free Chlorine

Cl₂

- 1** Fill DipCell to the fill line with sample.
- 2** Add 1 *Chlorine DPD #R Tablet (6999A). Cap and invert to mix.
- 3** Match sample color to a color standard.
Record result as ppm Free Chlorine.

*Potential health hazard. Read MSDS at www.lamotte.com

Total Chlorine

Cl₂

- 1** Remove cap. Add 1 *Chlorine DPD #3R tablet (6905A)
- 2** Cap DipCell and invert to mix.
- 3** Match sample color to a color standard.
Record result as ppm Total Chlorine.

NOTE: Total Chlorine - Free Chlorine = Combined Chlorine

*Potential health hazard. Read MSDS at www.lamotte.com

pH

- 1** If DipCell is empty, fill to line with sample.
- 2** Add 5 drops of *pH Indicator (7026). Cap and invert to mix.
- 3** Match sample color to a color standard.
Record as pH.